DEVICE CAP-10 FOR IN-PROCESS MEASUREMENT OF THE SINGLE-CORE WIRE CAPACITANCE PER UNIT LENGTH

ALEKSANDR GOLDSTEYN, GALINA VAVILOVA Department of Physical Methods of Non-Destructive Testing National Research Tomsk Polytechnic University Tomsk, Lenin Avenue 30 RUSSIA wgw@tpu.ru

Abstract: The paper describes an electrical-capacitive method to perform in-process measurement the capacitance of a single-core electric wire per unit length during its insulating. The optimum design values of the electro-capacitive measuring transducer are selected to minimize the impact of the design parametrs, and insulation and water properties. The block diagram of the device CAP-10 developed for implementation of the proposed method is presented. An algorithm for digital conversion of the output signal of the electro-capacitive measuring transducer enables correction of the effect of the change in water conductivity. Technical implementation of the device is described.

Key-Words: - single-core wire, insulation, capacitance per unit length, electro-capacitive measuring transducer, water conductivity, device.

1 Introduction

The capacitance stability along the entire wire length ensures the quality of information transmission in the cable line. Testing of the wire capacitance performed at the stage of wire insulation is most effective. In [1], the method of capacitance testing implies creation of the alternate field between the grounded wire core and the insulation surface via the electricalcapacitive measuring transducer (ECMT) immersed in the extrusion cooling bath. During in-process testing, the capacitance value of the capacitor is measured. This capacitor consists of a wire core as the first electrical component, and the cooling water in which the wire is immersed as the second electrical component. The water fills the space between the wire insulation surface and the ECMT with the test wire inside and provides electrical contact. The described test method is widely used in developments of the leading manufactures Sikora, Zumbach working in cable industry [2-51.

2 Design ECMT

Fig. 1 shows the design of the ECMT used for implementation of the described method. The ECMT consists of tubular measuring electrode 1 and a pair of tubular guard electrodes 2, 3 located inside a metal housing 4 and isolated from it by air layer 5. Single-core electric wire 6 continuously moves inside the ECMT. The guard electrodes are connected to alternating voltage generator 7. The wire core and the ECMT housing are earthed.

The basic ECMT design parameters are the length of the measuring and guard electrodes, the distance (gap) between these electrodes and the inner diameter of the electrodes. Computer simulation of the interaction of the ECMT electric field with the electric wire in COMSOL Multiphysics enables determination of the ECMT optimal design parameters to minimize the systematic error of the wire capacitance measurement [6].

Fig. 2 shows the distribution of the electric potentials for interaction of the ECMT electric

field with the tested wire plotted in equipotential lines and shades of gray for the case of guard electrodes of R radius, 1R length and 1 mm gap.



Fig. 2 Electric field created by the ECMT with guard electrodes with a length of 1R and a gap of 1 mm.

In this case, the electric field spreads at the distant edges of the guard electrodes that provides highly uniform longitudinal electric field in the measurement zone.

In [6] it is shown that guard electrodes with a length of 0.5R and the gap of (1...3) mm provide minimum dependence of the capacitance measurement result on the impact of the wire design parameters, insulation properties and changes in water conductivity.

3 Block Diagram for CAP-10

Fig. 3 shows the block diagram for CAP-10 which consists of alternating $U_m \sin \omega t$ voltage generator 1, ECMT 2, an analogue conversion unit 6, adapter 7, digital processing unit 8, and visualization unit 9. The ECMT output signal is current I_x , whose amplitude is proportional to the measured capacitance of the tested wire. This signal is applied to input analogue conversion unit 6, which includes current transformer (CT) 3, current-to-voltage converter (CVC) 4 and amplitude-phase detector (APD) 5.



Fig. 3 Block diagram of CAP-10: alternating voltage generator (1), ECMT (2), current transformer (3), current-to-voltage converter (4), amplitude-phase detector (5), analogue conversion unit (6), adapter (7), digital processing unit (8), visualization unit (9).



Fig. 1 ECMT design: measuring electrode (1), guard electrodes (2, 3), transducer housing (4), air layer (5), wire (6), alternating voltage generator (7).

CT is used for galvanic isolation of the electrode measuring circuit and the signal analogue conversion circuit. The CT output current I_x is applied to CVC, where it is converted into voltage, and then to the APD input, in which the real and imaginary parts of the signal measurement information are differentiated. The APD reference signal is the voltage of generator 1.

The APD output voltages (Re U μ Im U) proportional to the amplitude values of the complex parts of the current in the measuring electrode circuit are applied to adapter inputs 7.

The adapter is a data acquisition board which is used as a USB3000 module, a universal highspeed eight-channel ADC. The adapter converts the analogue signal into a digital code and transmits it to the personal computer (PC). Digital processing unit 8 and visualization unit 9 are implemented on PC using LabView 8.5 software.

Digital processing unit 8 performs filtering, measurement data averaging within 1 second and digital processing. Visualization block unit 9 provides display of the measurement results in the form convenient for an operator.

4 Algorithm for Digital Conversion of the Measuring Signal

The current value I_x (Fig. 1) in the measuring electrode circuit is linearly related to the capacitance per unit length C_w of the tested wire area:

$$C_{\rm w} = C_0 + k \cdot I_{\rm r}, \qquad (1)$$

where C_0 is a constant component, k is a proportionality factor whose values depend on the amplitude and frequency of voltage generator 7 and ECMT 2 design parameters. Values C_0 and k are experimentally determined during the initial adjustment of the CAP-10.

The water conductivity varies due to changes in the concentration and composition of the salts, acids and bases of the cooling water, as well as changes in water temperature. In [1], it is shown that the wire capacitance significantly depends on the electrical conductivity of the cooling water.

Without allowance for the actual water conductivity, the measurement error for the

capacitance per unit length can reach 20% or more depending on the range of the water electrical conductivity variation.

The offset from the impact of changes in water conductivity implies replacing the coefficients C_0 and k in conversion function (1) corresponding to the actual water conductivity (for example, when using distilled water with the saline concentration $\lambda \rightarrow 0$) with their corrected values $C_{01}(t)$ and $k_1(t)$.

The coefficients $C_{01}(t)$ and $k_1(t)$ are found based on indirect measurement of water conductivity. The ratio $t = \operatorname{Re} i / \operatorname{Im} i$ depends on the actual water conductivity and indicates the change in the phase angle φ between the vector current \dot{I}_x and the imaginary axis of the complex plane. Both of the functions can be approximated by the second order polynomials whose coefficients are found experimentally during the initial adjustment of the CAP-10.

5 Technical Implementation of CAP- 10

Physical configuration of the CAP-10 is shown in Fig. 4. The main control program of the CAP-10 is Cmetr.vi implemented via LabView 8.5. The program Cmetr.vi performs filtration, averaging within 1 second and digital processing measuring signal. Digital of processing of the measuring signal is conducted according to the algorithm for digital conversion of the measuring signal. The measurement result is obtained through the conversion function based on the evaluation of the measuring electrode current phase.



Fig. 4 Physical configuration of the CAP-10.

The front panel (Fig. 5) displays the following information:

• the complex parts of the measured current " $\operatorname{Re} \dot{I}_x$ " $\operatorname{Im} \dot{I}_x$ ";

• the output voltage and current of the generator " U_g " and " I_g ";

• the calculated amplitude values for the measured current " I_x " and its phase "Phase";

• the waveform graph for the wire capacitance per unit length "Actual value";

• the calculated value of the wire capacitance per unit length " C_w ";

• the button "STOP" to stop measurement;

• the waveform graph "Test log" to record the data array of the capacitance per unit length within a specified time interval;

• the button "Data recording" to start and stop test log file.

The performance of the CAP-10 and the efficiency of the proposed technique to offset from impact of changes in water conductivity on the results of measuring wire capacitance were assessed experimentally.



Fig.5 Front panel of the CAP-10.

The experimental data was obtained for single-core electric wires with the capacity ranging from 160°pF/m to 460°pF/m and saline concentration of water varying in the range of $(0...4)^{\circ}$ g/l.

The real values of the capacitance of the single-core wire per unit length were determined in accordance with GOST 27893-88 [7] using fresh water at room temperature $+(22\pm1)^{\circ}C$ and saline concentration close to zero $\lambda \rightarrow 0$. The studies obtained are shown in Fig. 6.

Figure 6 shows the range of the measured values for capacitance per unit length C_x under variation of water conductivity for different wires (with different real values of the capacitance per unit length C_r) without offset from the impact of changes in water conductivity (dotted lines), and for the case of the offset (solid lines).





The analysis of these results shows that without offset from the impact of changes in water conductivity the relative measurement error can reach 20 %, however, when offsetting, the values of the capacitance per unit length calculated by formula (1) for these ranges of parameters affecting the actual values differ by not more than 2.5 %. A limited range of changes affecting the parameters allows manyfold reduction in the measurement error.

6 Technical Characteristics of the CAP-10

The CAP is designed to measure the values of capacitance per unit length of the single-core wire, and to record and store the array of measurement data.

The main technical characteristics device are as follows:

- the test object is single-core insulated wire;
- the range of the measured capacitance per unit length is (50...750) pF/m;

• the outer diameter of the wire insulation varies from 0.5 mm to 12 mm;

• the maximum permissible measurement error for the capacitance per unit length is 2.5% from its nominal value in the range of the change in water conductivity equivalent to the variation in the NaCl concentration range $(0 \dots 4)$ g/l.

Conclusions

Thus, the device CAP-10 for in-process measurement of the single-core wire capacitance per unit length was developed on the basis of the ECMT immersed in the extrusion cooling bath. The values of the basic design parameters for ECMT were selected.

The design of the ECMT, the block diagram of the CAP-10 and the algorithm for measuring signal conversion are presented.

The conversion function is corrected based on indirect measurement of water conductivity. The techniques to offset from the impact of changes in water conductivity on the test result are suggested.

Significant decrease in the error in measurement of the capacitance per unit length of the wire proves the efficiency of the suggested technique.

References:

- Goldstein A E and Vavilova G V, Technological control of linear capacity of electric cables when major changes in water salinity, *Testing. Diagnostics*, No. 9, 2013, pp. 57-60.
- [2] Blohm W, Advanced on-line measuring and analysis techniques for an economical cable production, *Proceedings Intl. IWMA Conf. Economical Processing of Rod to Wire & Cable*, 1999, pp. 67-75.
- [3] CAPAC[®]/FFT, Capacitance measurement systems for cable production lines, Zumbach Electronic AG, SCHWEIZ: Hauptsitz, URL:http://www.zumbach.com/pdf/Litera ture_DE/Catalogs/CAPAC/CAPAC_Fami lie_CAPA.002.0002.D.pdf.
- [4] Fleming P and Coleman L R, A system for monitoring fluctuations in the thickness of a cable insulating sheath, 2001, GB Patent 2358928.
- [5] Fleming P and Coleman L, Capacitance monitoring systems, 2003, US Patent 20030128038.
- [6] Goldshtein A E, Vavilova G V and Belyankov V Yu, An electro-capacitive measuring transducer for the process inspection of the cable capacitance per

unit length in the process of production, *Russian Journal of Nondestructive Testing*, Vol. 51, No. 2, 2015, pp. 35-43/

[7] GOST (State Standard) 27893-88 Communication Cables. Methods of Tests, Moscow: Standards publishing house, 1989.